

Technical Bulletin #2

How To Replace Connector Fittings

This bulletin is intended to provide instruction for the replacement of “female connector fittings” on your HTP-1500 heat therapy system. This instruction can also be used with other heat therapy systems not manufactured by Adroit Medical systems, Inc.

First, remove the existing female connectors using scissors or a knife. Insert the replacement connectors (Part No. A2.2). That’s all. No glue or tension-rings are needed.

